

# TEST - 090213C FUEL - ULSD with...

PCS Instruments - HFRPC v2.10, Instrument D1021

UNIQUE TEST ID  
D1021-200902130749

## Fuel Information

Fuel Type Unknown fuel sample  
Fuel Name ULSD with...  
Fuel S/N 09-03613


## Test Information

Test Type ASTM D6079 (60degC)  
Test Avgs Temp 60.0degC Film 94% Fric 0.155  
Operator Name ja  
Upper Spec Batch Unnamed Batch  
Lower Spec Batch Unnamed Batch  
Time Test Started Friday, February 13, 2009 7:49 AM

## Wear Scar Information

Wear Scar X 233 um  
Wear Scar Y 151 um  
Wear Scar Avg 192.0 um  
Start Temp/RH 21.3 / 34.2  
Finish Temp/RH 22.1 / 34.6

Comments ULSD WITH 2 % BIODIESEL PLUS 1 mL OXY



Microscope wear scar measurement 1 of 1  
measured by JA on 2/13/2009 at 9:45 AM

