

TEST - 090211E FUEL - DSL

PCS Instruments - HFRPC v2.10, Instrument D1021

UNIQUE TEST ID
D1021-200902111452

Fuel Information

Fuel Type Unknown fuel sample
Fuel Name DSL
Fuel S/N 09-03612

Test Information

Test Type ASTM D6079 (60degC)
Test Avgs Temp 60.0degC Film 90% Fric 0.151

Wear Scar Information

Wear Scar X 242 um
Wear Scar Y 200 um
Wear Scar Avg 221.0 um

Additive Name PPM

Unnamed Additive

S/N

ULSD W/ 2% BIODIESEL

Operator Name

BW

Upper Spec Batch Unnamed Batch

Lower Spec Batch Unnamed Batch

Time Test Started

BW


Unnamed Batch

Unnamed Batch

Wednesday, February 11, 2009 2:52 PM

Start Temp/RH 23.8 / 37.2

Finish Temp/RH 23.7 / 35.2



Microscope wear scar measurement 1 of 1
measured by LC on 2/11/2009 at 4:21 PM

